

株式会社 オーヒラ

〒949-7135 新潟県南魚沼市新堀新田629-961

OHIRA Co., Ltd.

629-961 Niiborishinden,

Minamiuonuma-Shi, Niigata-Ken 949-7135 Japan

TEL 025-775-3891 FAX 025-775-3890



EC DECLARATION OF CONFORMITY

(No.DC-S02-C)

Manufacturer: Ohira Co., Ltd.

European Representative: Fausto Guzzetti
(Shin-Nippon Technical Center)
Via Solferino, 5 21057 Olgiate Olona(VA), Italy

Product: Trade Name: SLIT LAMP SL-102
Description: Slit Lamp

Classification: Class I without measuring function
Rule 12 according to Annex IX of the MDD

Conformity Assessment Route: Annex VII applied

We herewith declare that the above-mentioned products meet the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

General applicable directives: Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC
(MDD 93/42/EEC)

Standards: Harmonized Standards (published in the Official Journal of the European Communities) applicable to this product are:

EN ISO 15004-1:2009	Ophthalmic instruments - Fundamental requirements and test methods - Part 1: General requirements applicable to all ophthalmic instruments (ISO 15004-1:2006)
EN 60601-1:1990 +A1:1993+A2:1995	Medical electrical equipment - Part 1: General requirements for safety (IEC 60601-1:1988+A1:1991+A2:1995)
EN 60601-1-2:2007	Medical electrical equipment - Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests (IEC 60601-1-2:2007)
EN ISO 10993-1:2009	Biological evaluation of medical devices - Part 1: Evaluation and testing (ISO 10993-1:2003)

Signature:


Name Hiroji Ohira
Position President

Date: 2010.03.21